

The automatic calcimeter must be placed on a flat, horizontal and clean work surface, under a hood or in a well ventilated area.

It must not be subjected to vibration or shock.

For optimum performance, the temperature should be maintained at $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$.

The carousel must be perfectly positioned on the turntable in order to not slip during rotation.

The electronic scales is connected to calcimeter by RS232 cable.